Distributed by: **CLIAwaived.com™**
San Diego, CA 92121
tel 858-481-5031
toll free 888-882-7739
www.cliawaived.com

Procedure Card for Drugs of Abuse Cup

(Cups available with adulteration)

For detailed instructions please read the package insert. Store at 2-30°C (36-86°F) in the original sealed pouch. Do not freeze.

CUP COMPONENTS



ALSO INCLUDED:

- Package insert
- Security seal (if applicable)
- Adulteration color chart (if applicable)

COLLECTION & TESTING PROCEDURE

1 Collect the specimen. Sample must meet minimum urine level as indicated on the cup.



2 Secure test device cap to the specimen cup. NOTE: Screw the cap tightly to ensure adequate seal.



IMPORTANT: Cup lid must be secured tightly by twisting lid a quarter turn AFTER lid is snug

3 Turn cup on its side to activate testing. Place on a flat surface.



4 When applicable, read adulteration results by visually comparing color of reagent pads to corresponding blocks on the Color Chart at the time indicated.



5 Read results at 5 minutes. Do not interpret result after 10 minutes.



CLIAwaived™.com

ALL POSITIVE RESULTS ARE PRESUMPTIVE AND SHOULD BE CONFIRMED BY AN ALTERNATE METHOD (e.g. GC/MS)