
MATERIAL SAFETY DATA SHEET

PROADVANTAGE Vinyl Exam Gloves

Powder Free Product Numbers: P359401, P359402, P359403, P359404, P359405
Powdered Product Numbers: P359501, P359502, P359503, P359504, P359505

Physical Data – Composition

Polyvinyl Chloride	46.84%
Di-Octyl Phthalate (DOP)	44.10%
Calcium-Zinc Stabilizer	1.17%
D80	7.89%
Yellow Pigment	N/A

Fire and Explosions

Stability: Stable
Extinguishing Media: Foam, Carbon Dioxide, Water
Flash Point: N/A

Normal Precautions

Ventilation: Not Necessary Under Normal Conditions
Spillage Handling: N/A
Eyes: N/A
Storage: Store at Room Temperature

Precaution

This product contains polyvinyl chloride (PVC). If user or a patient are allergic to vinyl or PVC or experience any discomfort, discontinue use immediately and consult with a physician.

Do not reuse gloves. Good quality exam gloves provide an excellent biological barrier. They are not intended for use as a chemical barrier.

Storage

Store in original packing in a cool, dry and well-ventilated area, away from dust, sunlight, moisture, X-ray and excessive heat about 100° F (37° C)